

PHOTOMULTIPLICATEUR (PM NOIR)

- 34 mm useful diameter head-on type
- flat window
- semi-transparent bi-alkaline photocathode
- high stability
- good linearity
- for laboratory and industrial photometry
- XP2012 has a 14-pin all-glass base; XP2012B has a 12-pin plastic base

QUICK REFERENCE DATA

Radiant sensitivity characteristic	bi-alkaline
Useful diameter of the photocathode	> 34 mm
Cathode blue sensitivity	11,5 μ A/lmF
Supply voltage for anode blue sensitivity = 7,5 A/lmF	1350 V
Pulse amplitude resolution for ^{55}Fe	\approx 42%
Mean anode sensitivity deviation	\approx 1%
Anode pulse rise time	\approx 2,5 ns
Linearity (with voltage divider B)	up to \approx 200 mA

To be read in conjunction with *General Operational Recommendations Photomultiplier Tubes*.

GENERAL CHARACTERISTICS

notes

Window	
Material	lime glass
Shape	plano-plano
Refractive index at 400 nm	1,54
Photocathode	1
Semi-transparent, head-on	
Material	bi-alkaline
Useful diameter	> 34 mm
Radiant sensitivity characteristic	see Fig. 6
Maximum radiant sensitivity	400 ± 30 nm
Luminous sensitivity	$\approx 70 \mu\text{A/lm}$
Blue sensitivity	typ. 11,5 $\mu\text{A/lmF}$ $> 10 \mu\text{A/lmF}$
Radiant sensitivity at 400 nm	$\approx 90 \text{ mA/W}$

Multiplier system

Number of stages	10
Dynode structure	linear focused
Dynode material	Cu Be
Capacitances	
Anode to all	$\approx 5 \text{ pF}$
Anode to final dynode	$\approx 3 \text{ pF}$

Magnetic field

When the photocathode is illuminated uniformly the anode current is halved (at $V_{ht} = 1200 \text{ V}$, voltage divider A):

- at a magnetic flux density of 0,6 mT in the direction of the longitudinal axis;
- at a magnetic flux density of 0,35 mT perpendicular to axis a (see Fig.1);
- at a magnetic flux density of 0,15 mT parallel to axis a.

It is recommended that the tube be screened against the influence of magnetic fields by a mu-metal shield protruding > 15 mm beyond the photocathode.




Fig. 1 Axis a with respect to base pins (bottom view).

RECOMMENDED CIRCUITS




Fig. 2 Voltage divider A.

For optimum peak amplitude resolution it is recommended that the voltage between the first dynode and the photocathode be maintained at ≈ 200 V, e.g. by means of a voltage regulator diode.




Fig. 3 Voltage divider B.

Typical values of capacitors: 10 nF

- k = cathode
- dn = dynode no
- a = anode
- R_L = load resistor

TYPICAL CHARACTERISTICS			notes
With voltage divider A (Fig. 2)			2
Supply voltage for an anode blue sensitivity of 7,5 A/ImF (Fig. 8)	< typ.	1600 V 1350 V	i
Gain at $V_{ht} = 1350$ V (Fig. 9)	\approx	$6,5 \times 10^5$	
Anode dark current at an anode blue sensitivity of 7,5 A/ImF (Fig. 8)	< typ.	10 nA 1 nA	3,4
Pulse amplitude resolution for ^{137}Cs at an anode blue sensitivity of 1,5 A/ImF	\approx	7,2 %	5
Pulse amplitude resolution for ^{55}Fe at an anode blue sensitivity of 7,5 A/ImF	\approx	42 %	6
Peak-to-valley ratio for ^{55}Fe at an anode blue sensitivity of 7,5 A/ImF	\approx	34	6
Mean anode sensitivity deviation			13
long term (16 h)	\approx	1 %	
after change of count rate	\approx	1 %	
versus temperature between 0 and + 40 °C at 450 nm	\approx	0,2 %/K	
Anode current linear within 2% at $V_{ht} = 1350$ V	up to	\approx	65 mA
With voltage divider B (Fig. 3)			2
Anode blue sensitivity at $V_{ht} = 1700$ V (Fig. 8)	\approx	6,5 A/ImF	
Anode pulse rise time at $V_{ht} = 1700$ V	\approx	2,5 ns	7
Anode pulse duration at half height at $V_{ht} = 1700$ V	\approx	6 ns	7
Signal transit time at $V_{ht} = 1700$ V	\approx	26 ns	7
Anode current linear within 2% at $V_{ht} = 1700$ V	up to	\approx	200 mA
LIMITING VALUES (Absolute maximum rating system)			
Supply voltage	max.	1800 V	8
Continuous anode current	max.	0,2 mA	9
Voltage between first dynode and photocathode	max.	500 V	10
	min.	100 V	
Voltage between consecutive dynodes	max.	300 V	
Voltage between anode and final dynode	max.	300 V	11
	min.	30 V	
Ambient temperature range	max.	+ 80 °C	
Operational (for short periods of time)	min.	-30 °C	12
Continuous operating and storage	max.	+ 50 °C	
	min.	-30 °C	

Notes

1. The bialkaline photocathode has a significant resistance which increases rapidly with reducing temperature. It is thus recommended that it should not be subjected to light of too great an intensity; the cathode current should be limited, for example, to 1 nA at room temperature or 0,1 nA at -30 °C. If too high a photocurrent is passed, the cathode can no longer be considered to be an equipotential surface, and the focusing of electrons onto the first dynode will be affected, resulting in departures of linearity.
2. To obtain a peak pulse current greater than that obtainable with divider A, it is necessary to increase the inter-dynode voltage progressively. Divider circuit B is an example of a progressive divider, giving a compromise between gain, speed and linearity. Other dividers can be conceived to achieve other compromises. It is generally recommended that the voltage ratio between two successive stages is less than 2.
3. Wherever possible, the power supply should be arranged so that the cathode is earthed and the anode is at +HT. However, it is sometimes necessary to supply the tube with the anode earthed and the cathode at -HT; under these circumstances, noise and dark current will generally increase and become erratic, particularly after application of voltage. The glass envelope of the tube should be supported only by insulators with an insulating resistance greater than $10^{15} \Omega$. If a metal shield is used, it should be kept at the cathode potential.
4. Dark current is measured at ambient temperature, after the tube has been in darkness for approximately 1 min. Lower value can be obtained after a longer stabilisation period in darkness (approx. 30 min.).
5. Pulse amplitude resolution for ^{137}Cs is measured with an NaI (Tl) cylindrical scintillator with a diameter of 32 mm and a height of 32 mm. The count rate used is $\approx 10^3$ c/s.
6. Pulse amplitude resolution for ^{55}Fe is measured with an NaI (Tl) cylindrical scintillator with a diameter of 25 mm and a height of 1 mm provided with a beryllium window. The count rate used is 2×10^3 c/s.
7. Measured with a pulsed-light source, with a pulse duration (FWHM) of < 1 ns, the cathode being completely illuminated. The rise time is determined between 10% and 90% of the amplitude of the anode pulse. The signal transit time is measured between the instant at which the illuminating pulse at the cathode becomes maximum, and the instant at which the anode pulse reaches its maximum. Rise time, pulse duration and transit time vary as a function of high tension supply voltage V_{ht} , approximately as $V_{ht}^{-\frac{1}{2}}$.
8. Or the voltage at which the tube has an anode spectral sensitivity of 75 A/lmF (voltage given on test certificate for an anode blue sensitivity of 7,5 A/lmF, multiplied by 1,4), whichever is the lower.
9. A value of < 10 μA is recommended for applications requiring high stability.
10. Minimum value to obtain good collection in the input optics.
11. When calculating the anode voltage the voltage drop across the load resistor should be taken into account.
12. For types with plastic base this range of temperatures is limited principally by stresses in the sealing layer of the base to glass bulb.
13. The mean pulse amplitude deviation is measured by coupling an NaI (Tl) scintillator to the window of the tube. Long term (16 h) deviation is measured by placing a ^{137}Cs source at a distance from the scintillator such that the count rate is $\approx 10^4$ c/s corresponding to an anode current of ≈ 300 nA.
Mean pulse amplitude deviation after change of count rate is measured with a ^{137}Cs source at a distance of the scintillator such that the count rate can be changed from 10^4 c/s to 10^3 c/s corresponding to an anode current of $\approx 1 \mu\text{A}$ and $\approx 0,1 \mu\text{A}$ respectively.
Both tests are carried out according to ANSI-N42-9-1972 of IEEE recommendations.

MECHANICAL DATA

Dimensions in mm





Fig. 4 XP2012

XP2012B


Base: 14-pin all-glass

Base: 12-pin (JEDEC B12-43)


Net mass: 54 g

Net mass: 72 g

PIN CONNECTIONS



XP2012



XP2012B

ACCESSORIES

Socket

for XP2012 type FE1112
for XP2012B type FE1012

Mu-metal shield type 56609




Fig. 6 Radiant sensitivity characteristic.




Fig. 7 Relative gain as a function of the voltage between d5 and d4, normalized to V_d ; $V_{d6/d4}$ constant.

Note: Gain regulation by changing the voltage between d5 and d4 may cause a degradation of other parameters such as stability and linearity.




Fig. 8 Anode blue sensitivity, sa_F , and anode dark current, i_{da} , as a function of supply voltage V_{ht} .




Fig. 9 Gain G as a function of supply voltage V_{ht} .

X
RTC 1477

N76

photomultiplicateurs

⇒ (PM Noir)

Diam. utile de photo- cathode min. (mm)	Type	Nombre d'étages	Réponse spectrale (courbe page 16)	Sensibilité de cathode Val. moy.		Sensibilité anodique ou gain			Courant d'obscurité	
				En lumière blanche ($\mu\text{A} \cdot \text{lm}^{-1}$)	En lumière monochr. (mA, W^{-1}) (1)	Sensibilité $\text{A} \cdot \text{Im}^{-1}$ ou $\text{kA} \cdot \text{W}^{-1}$ (1)	Gain	Pour tension (val. moy.) (V)	Val. moy. (nA)	Pour sensibilité ($\text{A} \cdot \text{Im}^{-1}$) Gain (10^3 ou HT (V)
14	PM 1910	10	A (S 11)	60	60	$30 \text{kA} \cdot \text{W}^{-1}$	10^7	1 400	1,5	$30 \text{kA} \cdot \text{W}^{-1}$
	PM 1912	12	D	75		0,2		1 700	2	10^7
	PM 1920	6	A (S 11)	60	60	30		800	2	0,2
	XP 1110	10	A (S 11)	60	60	0,2		1 400	1,5	30
	XP 1113	6	A (S 11)	60	1,6	10		820	2	0,2
	XP 1116	10	C (S 1)	20				1 650	$5 \cdot 10^3$	10
	XP 1117	9	T (S 20)	140	13	30		1 520	10	30
	XP 1118	10	U (S 13)	60	60	30		1 400	2	30
	PM 1980	10	A (S 11)	75	70	$30 \text{kA} \cdot \text{W}^{-1}$		1 400	2	$30 \text{kA} \cdot \text{W}^{-1}$
	PM 1982	11	D	80		$250 \text{kA} \cdot \text{W}^{-1}$		1 450	2,5	$250 \text{kA} \cdot \text{W}^{-1}$
32	150 CVP	10	C (S 1)	20	1,4	10	$60 \text{kA} \cdot \text{W}^{-1}$	1 600	$2 \cdot 10^3$	10
	150 UVP	10	U (S 13)	85	75	60		1 500	10	60
	PM 2012	10	D	77		60		1 350	1	$60 \text{kA} \cdot \text{V}$
	XP 1011	10	SUPER A	90	80	60		1 500	10	60
	XP 1016	10	T (S 20)	160	16	60		1 460	3	60
	XP 1017	10	S 20 R	210	6,5	60		1 470	2	60
	XP 2008	10	SUPER A	80	70	60		1 180	5	60
	XP 2010	10	SUPER A	90	80	60		1 180	2	60
	remplace ← le 150 UVP									
44	56 AVP	14	A (S 11)	60	60	$3 \cdot 10^7$	$10 \text{kA} \cdot \text{W}^{-1}$	1 800	20	3...
	56 CVP	10	C (S 1)	20	1,4	10		1 800	$4 \cdot 10^3$	10
	56 DVP	14	D	80		$3 \cdot 10^7$		1 900	6	3...
	56 DUVP	14	DU	80		$3 \cdot 10^7$		1 900	6	3...
	56 SBUVP	14	SB	20		$3 \cdot 10^7$		2 050	30	3...
	56 TVP	14	T (S 20)	150	15	$3 \cdot 10^7$		2 050	60	3...
	56 TUVP	14	TU	150	15	$3 \cdot 10^7$		2 050	60	3...
	PM 2232 (2)	12	D	80		$3 \cdot 10^7$		1 900	7	3...
	XP 1230	12	D	90		$3 \cdot 10^7$		2 300	7	3...
	XP 2020	12	D	85		$3 \cdot 10^7$		2 200	7	3...
	XP 2230 (2)	12	D	90		$3 \cdot 10^7$		2 300	7	3...
	XP 1002	10	T (S 20)	165	16	60		1 460	3	60
	XP 1003	10	TU	165	16	60		1 460	5	60
	XP 1004	10	U (S 13)	80	70	60		1 450	10	60
68	XP 2000 (3) (6) 10 P		D	90		$10 \text{kA} \cdot \text{W}^{-1}$		1 250	1	1500
	PM 2312 (2)	12	D		80	$3 \cdot 10^7$		2 300	15	3...
110	XP 2030 (3) (6) 10 P		D		105	$10 \text{kA} \cdot \text{W}^{-1}$		1 220	1	1500
	XP 2040 (4)	14	A (5)	70	70	$3 \cdot 10^7$	$10 \text{kA} \cdot \text{W}^{-1}$	2 000	200	3...
	XP 2041 (4)	14	D (5)	85		$3 \cdot 10^7$		2 200	30	3...
200	XP 2050	(6) 10 P	D	95		$10 \text{kA} \cdot \text{W}^{-1}$		1 220	2	1500
	60 DVP (7)	12	D		70	$3 \cdot 10^7$		3 000	6	3...
10 cm ²	PM 555 (4)	7	bialcaline SbNa ₂ K		45	10^6		6 200	10	10

(1) Longueur d'onde de mesure : T et TU : $\lambda = 698 \text{ nm}$
 C : $\lambda = 903 \text{ nm}$
 SB : $\lambda = 250 \text{ nm}$
 S 20 R : $\lambda = 858 \text{ nm}$
 A, super A et U : $\lambda = 437 \text{ nm}$
 D et DU : $\lambda = 401 \text{ nm}$
 SbNa₂K : $\lambda = 404 \text{ nm}$.

(2) Tubes avec pied en verre pouvant être fournis avec culot sous l'appellation
 XP B ou PM B.

- (3) Tubes avec culot pouvant être fournis sans culot sous l'appellation
- (4) Tube avec habillage.
- (5) Etendue jusqu'à 200 nm grâce à un verre transparent à l'ultraviolet.
- (6) P = Dynodes « persiennes ».
- (7) Tube ne pouvant être livré avec habillage sur demande (60 DVP).
- (8) Spécifiée pour chaque tube sur sa feuille de caractéristiques.
- (9) Sortie coaxiale 50 Ω par connecteur General Radio GR 874.
- (10) Support livré avec le tube.
- (11) Dans les conditions optimales de répartition de tensions.

Caractéristiques de rapidité		Valeurs à ne pas dépasser (limites absolues)		Dimensions			Nombre de broches	Type de support (voir page 22)	Type
Temps de tracé de montée (ns)	Ecart centre-bord (ns)	T.H.T. (V)	Courant continu d'anode (mA)	Long. max. (mm)	Long. max. sans les broches (mm)	Diam. max. verrière (mm)			
-1	2,5		1 900	0,2	105	88	20	FE 1004	PM 1910
	2,5		2 000	0,2	100	90	20	sortie par fils	PM 1912
	2		1 200	0,2	70	60	20	FE 1004	PM 1920
	3,5		1 900	0,2	105	88	20	FE 1004	XP 1110
	1,3,5		1 200	0,2	70	60	20	XP 1113	XP 1113
	3,5		1 900	0,01	105	91	20,5	FE 1004	XP 1116
	2,3,5		1 900	0,2	105	91	20,5	FE 1004	XP 1117
	3,5		1 900	0,2	105	91	20	FE 1004	XP 1118
-1	3		1 900	0,2	98	88	26,5	B 8700 67	PM 1980
	2,5		2 000	0,2	98	88	29		PM 1982
-1	3,5		1 800	0,02	127	114	39,5	FE 1002	150 CVP
	3,5		1 800	0,2	127	114	39,5	FE 1002	150 UVP
	3,5		1 800	0,2	127	114	39,5	FE 1002	PM 2012
	3,5		1 800	0,2	127	114	39,5	FE 1002	XP 1011
	3,5		1 800	0,2	127	114	39,5	FE 1002	XP 1016
	3,5		1 800	0,2	127	114	39,5	FE 1002	XP 1017
	3,5		1 800	0,2	127	114	39,5	FE 1002	XP 2008
	3,6		1 800	0,2	127	114	39,5	FE 1002	XP 2010
0 ⁷	2,1	0,5	2 500	0,2	192	175	52,5	FE 1003	56 AVP
	2,1		3 000	0,02	174	157	52,5	FE 1003	56 CVP
	2,1		2 500	0,2	192	175	53,5	FE 1003	56 DVP
	2,1		2 500	0,2	192	175	53,5	FE 1003	56 DUVP
	2,1		2 500	0,2	192	175	53,5	FE 1003	56 SBUVP
	2,1		2 750	0,2	192	175	53,5	FE 1003	56 TVP
	2,1		2 750	0,2	192	175	53,5	FE 1003	56 TUVP
	2,2		3 000	0,2	139	122	53,5	FE 2019	PM 2232
	1,6		3 000	0,2	145	128	53,5	FE 2003	XP 1230
	1,5		3 000	0,2	192	175	53,5	FE 1003	XP 2020
	1,6		3 000	0,2	145	128	53,5	FE 2003	XP 2230 (2)
	4		1 800	0,2	148	127	52,5	FE 1001	XP 1002
	4		1 800	0,2	148	128	53,5	FE 1001	XP 1003
	4		1 800	0,2	148	128	53,5	FE 1001	XP 1004
	9		2 000	0,2	148	129	52,5	FE 1001	XP 2000 (3)
10 ⁷	2,5	1	3 000	0,2	184	166	77,5	FE 2019	PM 2312
	0		2 000	0,2	159	140	77,5	FE 1001	XP 2030 (3)
10 ⁷	2	1	3 000	0,2	281 (4)	264 (4)	136,5 (4)	FE 1003	XP 2040 (4)
	2		3 000	0,2	281 (4)	264 (4)	136,5 (4)	FE 1003	XP 2041 (4)
	6		2 000	0,2	195	176	130	FE 1001	XP 2050
10 ⁷	2,1	2	3 700	0,2	318 (7)	301 (7)	231,5 (7)	FE 1003	60 DVP (7)
	1,5		(8)	0,05	170 (4)		112 (4)	(9)	(10)
5									PM 555 (4)

Definitions des sélections

Sélection 03 : Tubes sélectionnés pour faible bruit de fond. Sélections standards : 56 DVP/03 et 56 DUVP/03.

Sélection A : Tubes ayant une sensibilité anodique identique, pour des tensions d'alimentation les plus voisines possibles. Sélections standards : 56 DVP/A, 56 DUVP/A, XP 2230/A, XP 1230/A et XP 2020/A.

Les caractéristiques des photomultiplicateurs référencés PM sont celles de tubes en développement et n'impliquent pas qu'elles restent identiques pour les tubes de série mis en fabrication.